

Report Issue Date: 6/15/2017

Steve Green
AT Labs, a Unit of Assay Technology
250 DeBartolo Place
Suite 2525
Boardman, OH 44512

Participant ID# 100903

Dear Steve Green,

Please find your organization's Environmental Lead Proficiency Analytical Testing results for **ELPAT Round 99**. It is the participant's responsibility to thoroughly review results and to immediately contact the AIHA Proficiency Analytical Testing Programs in writing, if any errors are found in your report.

The proficiency demonstrated by the results of this ELPAT round is valid until the results of the retest round are available on August 15, 2017, if the participant chooses to enroll, or until September 15, 2017 when the next ELPAT report will be available. Unacceptable performance may be improved by correctly analyzing a set of retest samples. If you require a retest for the round, you may order one by completing the Retest Order Form available online at www.aihapat.org. The completed form and payment must be received by June 27, 2017. Refer to the PAT Programs Schedule located at www.aihapat.org for important retest round dates.

Please handle, store and analyze your PAT samples in the same manner as routine client samples. To submit results, visit the Proficiency Analytical Testing (PAT) page and click on the PAT Data Entry Portal: www.aihapat.org. **Always print and save the confirmation page** after submitting data via the PAT Data Entry Portal.

Participants shall not describe their proficiency status in a manner that implies accreditation, certification or variations thereof. PAT results pertain only to the participant organization at the location listed on this results report. AIHA PAT Programs makes every effort to ensure that individual participant results are kept confidential and are not made public. Round results are only released to the participant and those entities requiring this information for accreditation, regulatory and contract purposes. New participants are made aware of the arrangement in advance of participation and consent is sought prior to the release of records for participants. PAT reports may not be reproduced or distributed unless copied in its entirety.

Any enrolled participant that is unable to participate in a PT round must request an "Excused Absence" in order to not receive outliers and an unacceptable performance rating. This written request must be received before the PT round closes. Please note that an "Excused Absence" will not be granted for more than two consecutive rounds.

ELPAT Round 100 sample kits will be mailed to participants around August 1, 2017. An email will be sent out upon shipment of the samples. If you do not receive samples within fifteen (15) days after the ship date please contact the AIHA PAT Programs. Your organization's data will be due by 11:59pm ET on September 1, 2017.

Samples are generated, characterized, packaged, and shipped by RTI International, Research Triangle Park, NC under contract with AIHA Proficiency Analytical Testing Programs. Unless otherwise noted, sample homogeneity and stability criteria were satisfied for all samples.

I encourage you to contact me with any feedback, questions or if you wish to contest your results at aoler@aiha.org.

Sincerely,



Angela Oler, ASQ CQA
Director of Operations, AIHA PAT Programs, LLC

Environmental Lead Proficiency Analytical Testing Results

This document contains three sub-reports relating to ELPAT Round 99. The first report contains your organization's results listed per contaminant, per sample. The second report contains your past proficiency data for 2 and 4 rounds respectively (where applicable), and the final report contains summary results for all participants for ELPAT Round 99.

Testing Results for ELPAT Round 99

This part of your report contains your organization's results listed per contaminant, per sample.

Contaminant	Units	#	Result	Reference Value	Lower Limit	Upper Limit	z-Score	Rating
Dust Wipe	µg	1	160.7	160.5	115.4	205.6	0.0	A
	µg	2	35.5	42.2	28.4	56.0	-1.5	A
	µg	3	296.8	350.6	246.6	454.5	-1.6	A
	µg	4	170.8	191.5	144.0	239.0	-1.3	A

Statistical Analysis Interpretation Note:

Reference value is the mean of the reference group.

Lower limit = reference value - 3 standard deviations; Upper limit = reference value + 3 standard deviations

z-Score = (reported result - reference value)/standard deviation. Note: z-Scores are used to predict trends and to indicate how far a particular score is away from the mean.

A – Acceptable* Analysis; U - Unacceptable Analysis

Both the assigned values and acceptance limits are based on consensus of the reference group. *The acceptability of reported results is based on upper and lower acceptance limits. This is why a reported result may appear unacceptable according to z-Score, but be identified as acceptable.

Any non-participation or non-reporting of PAT data will result in unacceptable results (see PAT Programs Participation Policies, Section 2.1.6.2.).

Overall Performance Summary Concluding with 99

The following table contains overall proficiency results for 2 and 4 rounds respectively (where applicable). For more information in regard to the determination of proficiency, please visit www.aihapat.org.

Sample	Round	Round Performance	2 Rounds	2 Round %	4 Rounds	4 Round %	Proficiency Status
Dust	96	4/4					
	97	4/4					
	98	4/4					
	99	4/4	8/8	100	16/16	100	P

Interpretation Note:

The denominators represent the total number of samples analyzed.

The numerators represent the number of acceptable results.

P – Proficient; NP – Non-proficient; I – Indeterminate (not enough rounds to determine proficiency)

A participant is rated proficient for the applicable ELPAT matrix if the participant's performance meets any of the following: (1) In the last two rounds, all samples are analyzed and the results are 100% acceptable; or (2) Three fourths (75%) or more of the accumulated results over four rounds are acceptable. A participant is rated non-proficient for the applicable matrix if the participant's performance does not meet either of the proficiency categories mentioned above.

The following items are available in the [Environmental Lead Scheme Plan](#).

Procedures used to statistically analyze the data, establish any assigned value and standard deviation for proficiency assessment, or other criteria for evaluation; details of the metrological traceability and measurement uncertainty of any assigned value; information about design and implementation of PT scheme. Environmental Lead Scheme Plan available at <http://www.aihapat.org/documents-policies-fees/Pages/default.aspx>.

Measurement uncertainty of any assigned value is also available on the respective certificate of analysis for the round.

Technical Comment: No remarkable observations.

Performance of all Participants for ELPAT Round 99

The following table contains aggregate results for all participants for ELPAT Round 99.

Contaminant	#	Ref. Value	Ref. Std. Dev.	RSD (%)	Uncertainty Measurement	Total Participants	Total Acceptable	Low*	High*
Paint Chips	1	2.4810	0.1339	5.4	0.014787	143	137	4	2
	2	0.5736	0.0386	6.7	0.004264	143	138	2	3
	3	2.8571	0.1892	6.6	0.020891	143	136	4	3
	4	0.0780	0.0059	7.6	0.000652	143	139	2	2
Soil	1	221.5	14.7	6.6	1.580618	119	116	0	3
	2	57.5	6.3	11.0	0.684070	119	114	3	2
	3	171.7	12.5	7.3	1.345274	119	118	1	0
	4	326.3	21.7	6.7	2.340726	119	115	0	4
Dust Wipe	1	160.5	15.0	9.4	1.592461	130	125	3	2
	2	42.2	4.6	10.9	0.487804	130	126	2	2
	3	350.6	34.6	9.9	3.672760	130	126	3	1
	4	191.5	15.8	8.3	1.678849	130	125	4	1

Note:
 ***Low** - number of participant results that are less than the Lower Limit; ***High** - number of participant results that are greater than the Upper Limit.
 Reference group/participant data sets for individual methods are not separated out during statistical analysis. Methods used by participants produce comparable data based upon the proficiency samples provided. Methods represented by fewer than eight participant data points are not assessed for comparability.
 Additional technical comments or recommendations, when available, shall be shared with participants via the web and participants shall be notified via email.